

Support plane analysis

The support plane analysis is used for analysing [fitting](#) parameter error intervals. It works by calculating the reduced χ^2 along a 'cleverly' chosen path through the parameter space. For illustration let's start at the best fit parameter set, which can be regarded as a single point in the parameter space. Now we remove the parameter for which we want to calculate the error intervals from this location, that is, we take a single step parallel to its parameter axis. Now we adjust all the other parameters to "get down as far as possible" for this new situation. Usually this is done by keeping the removed parameter fixed and fitting all the other parameters. By this we deviate from the axis-parallel direction, in general. We then calculate the reduced χ^2 for this optimised parameter set. By iterating this procedure we get χ^2 as a function of the parameter of interest. The intersection points of this function with a given χ^2 confidence limit define the boundaries of the confidence interval of the parameter.

Advantages and disadvantages

Since [fitting](#) is used to derive a functional dependence of the reduced χ^2 minute convergence is of crucial importance, even more than for the [bootstrap method](#). If the convergence criterion is not strict enough, the χ^2 function suffers from numerical noise and is imprecise. Furthermore the support plane analysis is limited to [least squares](#) fitting and cannot be used with [MLE](#).

On the other hand the computational effort to calculate the confidence interval of a single parameter is less than for the [bootstrap method](#).

References

- Lakowicz JR (1999) *Principles of Fluorescence Spectroscopy*, 2nd edn. Kluwer Academic/Plenum Publishers, New York

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH
Rudower Chaussee 29 (IGZ)
12489 Berlin
Germany

P +49-(0)30-1208820-89
F +49-(0)30-1208820-90
info@picoquant.com
www.picoquant.com